

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040419\
 Data File : VX008664.D
 Acq On : 05 Apr 2019 02:59
 Operator : JC/SP
 Sample : K2274-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW042-190403

Quant Time: Apr 05 09:22:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	215364	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	360609	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	314523	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	131429	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	149661	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
35) Dibromofluoromethane	5.50	113	109534	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
50) Toluene-d8	8.71	98	450633	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
62) 4-Bromofluorobenzene	11.13	95	154104	47.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.68%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	53495	17.277	ug/l	90
44) Trichloroethene	7.21	130	16942	6.194	ug/l	97
64) Tetrachloroethene	9.34	164	6050	2.596	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040419\
Data File : VX008664.D
Acq On : 05 Apr 2019 02:59
Operator : JC/SP
Sample : K2274-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW042-190403

Quant Time: Apr 05 09:22:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

