

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX040419\
 Data File : VX008670.D
 Acq On : 05 Apr 2019 05:19
 Operator : JC/SP
 Sample : K2274-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST006MW148-190402

Quant Time: Apr 05 09:36:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	214869	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	355400	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	308452	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	130508	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	144932	46.25	ug/l	0.00
Spiked Amount			Recovery	=	92.50%	
35) Dibromofluoromethane	5.50	113	108657	47.82	ug/l	0.00
Spiked Amount			Recovery	=	95.64%	
50) Toluene-d8	8.71	98	447207	49.43	ug/l	0.00
Spiked Amount			Recovery	=	98.86%	
62) 4-Bromofluorobenzene	11.13	95	149918	46.73	ug/l	0.00
Spiked Amount			Recovery	=	93.46%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	148662	44.463	ug/l	97
11) Tert butyl alcohol	3.03	59	2526	3.443	ug/l #	72
16) Acetone	2.44	43	2164	1.199	ug/l	97
21) trans-1,2-Dichloroethene	3.17	96	2085	0.776	ug/l	90
24) 1,1-Dichloroethane	3.70	63	3738	0.689	ug/l #	89
27) cis-1,2-Dichloroethene	4.60	96	154114	49.888	ug/l	89
31) Cyclohexane	5.57	56	19097	3.527	ug/l	95
32) 1,1,1-Trichloroethane	5.49	97	3029	0.773	ug/l #	46
39) Methylcyclohexane	7.46	83	46719	9.735	ug/l #	75
40) Benzene	6.15	78	10540	0.911	ug/l	96
44) Trichloroethene	7.22	130	4238	1.572	ug/l	87
64) Tetrachloroethene	9.34	164	4769	2.087	ug/l	95
65) Chlorobenzene	10.13	112	5758	0.842	ug/l	96
67) Ethyl Benzene	10.25	91	6890	0.540	ug/l	100
68) m/p-Xylenes	10.36	106	1971	0.430	ug/l	82
73) Isopropylbenzene	11.02	105	4825	0.446	ug/l	99
83) tert-Butylbenzene	11.77	119	13255	1.518	ug/l	91
85) sec-Butylbenzene	11.94	105	25172	2.426	ug/l	96
89) n-Butylbenzene	12.38	91	5290	0.606	ug/l #	58
91) 1,2-Dichlorobenzene	12.39	146	1364	0.295	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040419\
Data File : VX008670.D
Acq On : 05 Apr 2019 05:19
Operator : JC/SP
Sample : K2274-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST006MW148-190402

Quant Time: Apr 05 09:36:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

