


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX040419\
Data File : VX008623.D
Acq On : 04 Apr 2019 09:00
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Title : SW846 8260
Last Update : Thu Apr 04 04:47:41 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	11796	PASS
75	95	30	60	51.6	29688	PASS
95	95	100	100	100.0	57482	PASS
96	95	5	9	6.5	3718	PASS
173	174	0.00	2	1.4	610	PASS
174	95	50	100	74.8	42968	PASS
175	174	5	9	7.0	3012	PASS
176	174	95	101	95.9	41208	PASS
177	176	5	9	6.6	2716	PASS