

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040422\
 Data File : VX027952.D
 Acq On : 04 Apr 2022 12:36
 Operator : JC/MD
 Sample : N2249-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Quant Time: Apr 05 05:48:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77692	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	127696	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	110547	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45246	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63129	62.975	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 125.960%	
35) Dibromofluoromethane	5.391	113	45329	54.994	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 109.980%	
50) Toluene-d8	8.653	98	150489	51.892	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.780%	
62) 4-Bromofluorobenzene	11.079	95	62908	51.044	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 102.080%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1694	3.818	ug/l	97
52) Toluene	8.720	92	758	0.383	ug/l	92
67) Ethyl Benzene	10.195	91	1907	0.483	ug/l	100
68) m/p-Xylenes	10.305	106	3726	2.564	ug/l	89
69) o-Xylene	10.646	106	1099	0.774	ug/l	99
73) Isopropylbenzene	10.963	105	1399	0.415	ug/l	88
80) 1,3,5-Trimethylbenzene	11.451	105	1035	0.354	ug/l	89
84) 1,2,4-Trimethylbenzene	11.756	105	5351	1.836	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040422\
Data File : VX027952.D
Acq On : 04 Apr 2022 12:36
Operator : JC/MD
Sample : N2249-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Quant Time: Apr 05 05:48:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

