

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040422\
 Data File : VX027964.D
 Acq On : 04 Apr 2022 17:44
 Operator : JC/MD
 Sample : PB143824ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB143824ZHE#01

Quant Time: Apr 05 05:52:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	92317	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	148850	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51054	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	47337	39.741	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	79.480%
35) Dibromofluoromethane	5.391	113	45218	47.063	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.120%
50) Toluene-d8	8.653	98	167518	49.554	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.100%
62) 4-Bromofluorobenzene	11.079	95	69214	48.180	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.360%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3451	6.546	ug/l #	84
20) Methylene Chloride	2.788	84	2915	3.074	ug/l	88
43) Isopropyl Acetate	6.342	43	5857	2.438	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040422\
Data File : VX027964.D
Acq On : 04 Apr 2022 17:44
Operator : JC/MD
Sample : PB143824ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB143824ZHE#01

Quant Time: Apr 05 05:52:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

