

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040422\
 Data File : VX027966.D
 Acq On : 04 Apr 2022 18:30
 Operator : JC/MD
 Sample : PB143824ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB143824ZHE#03

Quant Time: Apr 05 05:52:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	90045	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	149144	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53600	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59968	51.615	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.240%	
35) Dibromofluoromethane	5.385	113	47842	49.696	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.400%	
50) Toluene-d8	8.653	98	170099	50.219	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.440%	
62) 4-Bromofluorobenzene	11.085	95	71275	49.516	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 99.040%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	8194	15.936	ug/l	96
20) Methylene Chloride	2.794	84	65662	70.996	ug/l	90
43) Isopropyl Acetate	6.348	43	8033	3.337	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040422\
Data File : VX027966.D
Acq On : 04 Apr 2022 18:30
Operator : JC/MD
Sample : PB143824ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB143824ZHE#03

Quant Time: Apr 05 05:52:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

