

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040422\
 Data File : VX027968.D
 Acq On : 04 Apr 2022 19:17
 Operator : JC/MD
 Sample : PB143824ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB143824ZHE#05

Quant Time: Apr 05 05:53:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	93834	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	148479	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132766	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55292	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	56485	46.654	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 93.300%			
35) Dibromofluoromethane	5.391	113	48689	50.803	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.600%			
50) Toluene-d8	8.653	98	174873	51.859	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.720%			
62) 4-Bromofluorobenzene	11.079	95	73273	51.132	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 102.260%			
Target Compounds						
16) Acetone	2.380	43	8405	15.686	ug/l	94
20) Methylene Chloride	2.788	84	65799	68.272	ug/l #	87
43) Isopropyl Acetate	6.336	43	8257	3.445	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040422\
Data File : VX027968.D
Acq On : 04 Apr 2022 19:17
Operator : JC/MD
Sample : PB143824ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB143824ZHE#05

Quant Time: Apr 05 05:53:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

