

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040422\
 Data File : VX027972.D
 Acq On : 04 Apr 2022 20:50
 Operator : JC/MD
 Sample : N2229-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SOIL-DRUM

Quant Time: Apr 05 05:54:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	96943	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	158173	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	136176	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58900	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	52845	42.248	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	84.500%
35) Dibromofluoromethane	5.391	113	47468	46.493	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.980%
50) Toluene-d8	8.652	98	180950	50.373	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.740%
62) 4-Bromofluorobenzene	11.085	95	68807	45.073	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.140%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	4744	8.570	ug/l #	83
20) Methylene Chloride	2.788	84	2468	2.479	ug/l	88
43) Isopropyl Acetate	6.348	43	6942	2.719	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040422\
Data File : VX027972.D
Acq On : 04 Apr 2022 20:50
Operator : JC/MD
Sample : N2229-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SOIL-DRUM

Quant Time: Apr 05 05:54:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

