

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040422\
 Data File : VX027947.D
 Acq On : 04 Apr 2022 10:09
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 05 05:41:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	50.000	47.438	5.1	81	0.00
3 P	Chloromethane	50.000	40.491	19.0	76	0.00
4 C	Vinyl Chloride	50.000	42.359	15.3#	78	0.00
5 T	Bromomethane	50.000	41.851	16.3	85	0.00
6 T	Chloroethane	50.000	40.955	18.1	81	0.00
7 T	Trichlorofluoromethane	50.000	47.403	5.2	88	0.00
8 T	Diethyl Ether	50.000	47.368	5.3	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.220	5.6	89	0.00
10 T	Methyl Iodide	50.000	33.980	32.0#	62	0.00
11 T	Tert butyl alcohol	250.000	211.326	15.5	77	0.00
12 CM	1,1-Dichloroethene	50.000	42.528	14.9#	85	0.00
13 T	Acrolein	250.000	234.184	6.3	87	0.00
14 T	Allyl chloride	50.000	45.819	8.4	88	0.00
15 T	Acrylonitrile	250.000	233.713	6.5	85	0.00
16 T	Acetone	250.000	244.183	2.3	93	0.00
17 T	Carbon Disulfide	50.000	40.527	18.9	83	0.00
18 T	Methyl Acetate	50.000	42.529	14.9	81	0.00
19 T	Methyl tert-butyl Ether	50.000	48.584	2.8	90	0.00
20 T	Methylene Chloride	50.000	43.038	13.9	84	0.00
21 T	trans-1,2-Dichloroethene	50.000	41.507	17.0	82	0.00
22 T	Diisopropyl ether	50.000	49.399	1.2	103	0.00
23 T	Vinyl Acetate	250.000	240.506	3.8	98	0.00
24 P	1,1-Dichloroethane	50.000	47.562	4.9	98	0.00
25 T	2-Butanone	250.000	238.183	4.7	98	0.00
26 T	2,2-Dichloropropane	50.000	54.497	-9.0	110	0.00
27 T	cis-1,2-Dichloroethene	50.000	45.561	8.9	87	0.00
28 T	Bromochloromethane	50.000	48.362	3.3	100	0.00
29 T	Tetrahydrofuran	250.000	225.601	9.8	93	-0.01
30 C	Chloroform	50.000	48.658	2.7#	97	-0.01
31 T	Cyclohexane	50.000	41.980	16.0	83	0.00
32 T	1,1,1-Trichloroethane	50.000	48.368	3.3	96	0.00
33 S	1,2-Dichloroethane-d4	50.000	45.527	8.9	87	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	95	0.00
35 S	Dibromofluoromethane	50.000	50.437	-0.9	97	-0.01
36 T	1,1-Dichloropropene	50.000	46.771	6.5	87	0.00
37 T	Ethyl Acetate	50.000	46.655	6.7	95	0.00
38 T	Carbon Tetrachloride	50.000	54.925	-9.8	99	0.00
39 T	Methylcyclohexane	50.000	46.322	7.4	86	0.00
40 TM	Benzene	50.000	46.701	6.6	84	0.00
41 T	Methacrylonitrile	50.000	50.073	-0.1	96	0.00
42 TM	1,2-Dichloroethane	50.000	45.287	9.4	77	-0.01
43 T	Isopropyl Acetate	50.000	42.351	15.3	70	0.00
44 TM	Trichloroethene	50.000	49.008	2.0	92	0.00
45 C	1,2-Dichloropropane	50.000	48.416	3.2#	86	0.00
46 T	Dibromomethane	50.000	48.396	3.2	87	0.00
47 T	Bromodichloromethane	50.000	51.459	-2.9	92	0.00
48 T	Methyl methacrylate	50.000	42.216	15.6	73	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040422\
 Data File : VX027947.D
 Acq On : 04 Apr 2022 10:09
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 05 05:41:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 04:39:27 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	822.455	17.8	71	0.00
50 S	Toluene-d8	50.000	48.490	3.0	92	0.00
51 T	4-Methyl-2-Pentanone	250.000	222.606	11.0	77	0.00
52 CM	Toluene	50.000	47.623	4.8#	86	0.00
53 T	t-1,3-Dichloropropene	50.000	52.627	-5.3	92	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.051	-4.1	91	0.00
55 T	1,1,2-Trichloroethane	50.000	48.832	2.3	89	0.00
56 T	Ethyl methacrylate	50.000	47.740	4.5	83	0.00
57 T	1,3-Dichloropropane	50.000	47.786	4.4	86	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	242.356	3.1	83	0.00
59 T	2-Hexanone	250.000	227.595	9.0	78	0.00
60 T	Dibromochloromethane	50.000	56.531	-13.1	100	0.00
61 T	1,2-Dibromoethane	50.000	48.231	3.5	86	0.00
62 S	4-Bromofluorobenzene	50.000	48.426	3.1	91	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	93	0.00
64 T	Tetrachloroethene	50.000	47.788	4.4	87	0.00
65 PM	Chlorobenzene	50.000	47.763	4.5	88	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.964	-3.9	92	0.00
67 C	Ethyl Benzene	50.000	48.317	3.4#	86	0.00
68 T	m/p-Xylenes	100.000	96.550	3.5	86	0.00
69 T	o-Xylene	50.000	49.603	0.8	87	0.00
70 T	Styrene	50.000	50.528	-1.1	89	0.00
71 P	Bromoform	50.000	57.681	-15.4	110	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	100	0.00
73 T	Isopropylbenzene	50.000	45.674	8.7	89	0.00
74 T	N-amyl acetate	50.000	42.416	15.2	78	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	46.926	6.1	90	0.00
76 T	1,2,3-Trichloropropane	50.000	47.964	4.1	89	0.00
77 T	Bromobenzene	50.000	47.148	5.7	90	0.00
78 T	n-propylbenzene	50.000	45.650	8.7	90	0.00
79 T	2-Chlorotoluene	50.000	44.517	11.0	89	0.00
80 T	1,3,5-Trimethylbenzene	50.000	46.120	7.8	90	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.662	0.7	98	0.00
82 T	4-Chlorotoluene	50.000	45.688	8.6	89	0.00
83 T	tert-Butylbenzene	50.000	46.311	7.4	91	0.00
84 T	1,2,4-Trimethylbenzene	50.000	46.431	7.1	90	0.00
85 T	sec-Butylbenzene	50.000	47.973	4.1	92	0.00
86 T	p-Isopropyltoluene	50.000	47.710	4.6	91	0.00
87 T	1,3-Dichlorobenzene	50.000	47.463	5.1	95	0.00
88 T	1,4-Dichlorobenzene	50.000	46.232	7.5	94	0.00
89 T	n-Butylbenzene	50.000	49.505	1.0	94	0.00
90 T	Hexachloroethane	50.000	58.072	-16.1	113	0.00
91 T	1,2-Dichlorobenzene	50.000	47.436	5.1	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.833	4.3	89	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.146	1.7	99	0.00
94 T	Hexachlorobutadiene	50.000	52.260	-4.5	103	0.00
95 T	Naphthalene	50.000	48.029	3.9	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040422\
Data File : VX027947.D
Acq On : 04 Apr 2022 10:09
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 05 05:41:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 04:39:27 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.899	0.2	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6