



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040422\
 Data File : VX027946.D
 Acq On : 04 Apr 2022 09:34
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032922W.M
 Title : SW846 8260
 Last Update : Tue Apr 05 14:48:33 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	3873	PASS
75	95	30	60	54.9	11053	PASS
95	95	100	100	100.0	20120	PASS
96	95	5	9	6.7	1356	PASS
173	174	0.00	2	1.4	180	PASS
174	95	50	100	65.4	13162	PASS
175	174	5	9	7.9	1043	PASS
176	174	95	101	98.5	12958	PASS
177	176	5	9	6.6	854	PASS