

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040424\
 Data File : VX040883.D
 Acq On : 04 Apr 2024 13:18
 Operator : JC/MD
 Sample : P1974-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SOIL-2

Quant Time: Apr 05 04:10:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	43788	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	80242	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	83124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	42576	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	39040	49.690	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.380%
35) Dibromofluoromethane	5.379	113	27657	48.662	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.320%
50) Toluene-d8	8.647	98	102735	50.024	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.040%
62) 4-Bromofluorobenzene	11.079	95	43309	52.806	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.620%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	9807	32.147	ug/l	97
20) Methylene Chloride	2.794	84	2364	4.181	ug/l	96
43) Isopropyl Acetate	6.336	43	32420	21.170	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040424\
Data File : VX040883.D
Acq On : 04 Apr 2024 13:18
Operator : JC/MD
Sample : P1974-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SOIL-2

Quant Time: Apr 05 04:10:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

