

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX040424\  
 Data File : VX040879.D  
 Acq On : 04 Apr 2024 11:26  
 Operator : JC/MD  
 Sample : P1984-01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RSB-02-71-040324

Quant Time: Apr 05 04:08:33 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040124W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 02 05:40:55 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.550  | 168   | 42872    | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.757  | 114   | 75193    | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.049 | 117   | 76187    | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.018 | 152   | 39048    | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 37232    | 48.401   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 96.800%  |
| 35) Dibromofluoromethane    | 5.379  | 113   | 26384    | 49.539   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 99.080%  |
| 50) Toluene-d8              | 8.647  | 98    | 96201    | 49.987   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 99.980%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 39611    | 51.540   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 103.080% |


|                  |        |
|------------------|--------|
| Target Compounds | Qvalue |
| -----            |        |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX040424\  
Data File : VX040879.D  
Acq On : 04 Apr 2024 11:26  
Operator : JC/MD  
Sample : P1984-01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
RSB-02-71-040324

Quant Time: Apr 05 04:08:33 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040124W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 02 05:40:55 2024  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.550 min Scan# 71  
Delta R.T. -0.006 min  
Lab File: VX040879.D  
Acq: 04 Apr 2024 11:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
RSB-02-71-040324


Tgt Ion:168 Resp: 42872  
Ion Ratio Lower Upper  
168 100  
99 62.1 56.6 85.0



#33  
1,2-Dichloroethane-d4  
Concen: 48.401 ug/l  
RT: 5.952 min Scan# 798  
Delta R.T. 0.000 min  
Lab File: VX040879.D  
Acq: 04 Apr 2024 11:26

Tgt Ion: 65 Resp: 37232  
Ion Ratio Lower Upper  
65 100  
67 49.0 0.0 100.2









#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.757 min Scan# 91  
Delta R.T. -0.006 min  
Lab File: VX040879.D  
Acq: 04 Apr 2024 11:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
RSB-02-71-040324


Tgt Ion:114 Resp: 75193  
Ion Ratio Lower Upper  
114 100  
63 22.9 0.0 45.0  
88 16.2 0.0 36.6



#35  
Dibromofluoromethane  
Concen: 49.539 ug/l  
RT: 5.379 min Scan# 704  
Delta R.T. -0.006 min  
Lab File: VX040879.D  
Acq: 04 Apr 2024 11:26

Tgt Ion:113 Resp: 26384  
Ion Ratio Lower Upper  
113 100  
111 104.7 82.9 124.3  
192 17.9 13.3 19.9









#50  
Toluene-d8  
Concen: 49.987 ug/l  
RT: 8.647 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VX040879.D  
Acq: 04 Apr 2024 11:26




Instrument :  
MSVOA\_X  
ClientSampleId :  
RSB-02-71-040324


Tgt Ion: 98 Resp: 96201  
Ion Ratio Lower Upper  
98 100  
100 63.1 53.1 79.7



#62  
4-Bromofluorobenzene  
Concen: 51.540 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. 0.000 min  
Lab File: VX040879.D  
Acq: 04 Apr 2024 11:26

Tgt Ion: 95 Resp: 39611  
Ion Ratio Lower Upper  
95 100  
174 71.1 0.0 135.6  
176 68.7 0.0 131.0









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.049 min Scan# 1470  
Delta R.T. -0.006 min  
Lab File: VX040879.D  
Acq: 04 Apr 2024 11:26

Instrument :  
MSVOA\_X  
ClientSampleId :  
RSB-02-71-040324

Tgt Ion:117 Resp: 76187  
Ion Ratio Lower Upper  
117 100  
82 61.9 46.2 69.4  
119 30.6 25.5 38.3



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.018 min Scan# 1793  
Delta R.T. -0.006 min  
Lab File: VX040879.D  
Acq: 04 Apr 2024 11:26

Tgt Ion:152 Resp: 39048  
Ion Ratio Lower Upper  
152 100  
115 65.8 43.3 129.9  
150 158.2 0.0 345.0

