

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
 Data File : VX045594.D
 Acq On : 04 Apr 2025 13:13
 Operator : JC/MD
 Sample : Q1731-04 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-03B-90-040325

Quant Time: Apr 04 23:11:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	65889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	130842	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49298	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66193	54.935	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.860%	
35) Dibromofluoromethane	5.385	113	47755	51.442	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.880%	
50) Toluene-d8	8.647	98	163265	50.387	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.780%	
62) 4-Bromofluorobenzene	11.079	95	59950	50.794	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.580%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	13184	16.286	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	6377	6.477	ug/l	79
44) Trichloroethene	7.123	130	16052	17.642	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045594.D
Acq On : 04 Apr 2025 13:13
Operator : JC/MD
Sample : Q1731-04 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-03B-90-040325

Quant Time: Apr 04 23:11:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

