

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
 Data File : VX045601.D
 Acq On : 04 Apr 2025 15:56
 Operator : JC/MD
 Sample : PB167448TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167448TB

Quant Time: Apr 04 23:13:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	61250	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	119749	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111977	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	44818	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62436	55.741	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.480%
35) Dibromofluoromethane	5.379	113	43855	51.617	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.240%
50) Toluene-d8	8.647	98	153987	51.926	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.860%
62) 4-Bromofluorobenzene	11.079	95	54492	50.447	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.900%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	43321	94.187	ug/l	97
18) Methyl Acetate	2.709	43	1862	1.758	ug/l	94
20) Methylene Chloride	2.782	84	27654	32.116	ug/l	94
25) 2-Butanone	4.586	43	4003	5.946	ug/l	95
43) Isopropyl Acetate	6.348	43	7755	3.539	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045601.D
Acq On : 04 Apr 2025 15:56
Operator : JC/MD
Sample : PB167448TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167448TB

Quant Time: Apr 04 23:13:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

