

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
 Data File : VX045603.D
 Acq On : 04 Apr 2025 16:42
 Operator : JC/MD
 Sample : PB167448ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167448ZHE#02

Quant Time: Apr 04 23:13:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	62613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126809	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	120742	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48070	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	62443	54.534	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.060%	
35) Dibromofluoromethane	5.379	113	45031	50.051	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.100%	
50) Toluene-d8	8.647	98	161434	51.406	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.820%	
62) 4-Bromofluorobenzene	11.079	95	57704	50.446	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 100.900%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	40578	86.303	ug/l	97
18) Methyl Acetate	2.703	43	2603	2.405	ug/l #	82
20) Methylene Chloride	2.782	84	27153	30.847	ug/l	92
25) 2-Butanone	4.587	43	4670	6.785	ug/l	98
43) Isopropyl Acetate	6.354	43	7971	3.435	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045603.D
Acq On : 04 Apr 2025 16:42
Operator : JC/MD
Sample : PB167448ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167448ZHE#02

Quant Time: Apr 04 23:13:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

