

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
 Data File : VX045604.D
 Acq On : 04 Apr 2025 17:06
 Operator : JC/MD
 Sample : PB167448ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167448ZHE#03

Quant Time: Apr 04 23:13:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67347	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	136039	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132139	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54499	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	68468	55.593	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.180%
35) Dibromofluoromethane	5.385	113	49136	50.908	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.820%
50) Toluene-d8	8.647	98	175374	52.056	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.120%
62) 4-Bromofluorobenzene	11.079	95	65541	53.410	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.820%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	44052	87.106	ug/l	97
18) Methyl Acetate	2.709	43	2679	2.301	ug/l	91
20) Methylene Chloride	2.782	84	28336	29.929	ug/l	98
25) 2-Butanone	4.580	43	4787	6.466	ug/l	98
43) Isopropyl Acetate	6.348	43	8111	3.259	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045604.D
Acq On : 04 Apr 2025 17:06
Operator : JC/MD
Sample : PB167448ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167448ZHE#03

Quant Time: Apr 04 23:13:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

