

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
 Data File : VX045605.D
 Acq On : 04 Apr 2025 17:29
 Operator : JC/MD
 Sample : PB167448ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167448ZHE#04

Quant Time: Apr 04 23:13:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	68787	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	135303	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	130586	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52425	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68812	54.702	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.400%	
35) Dibromofluoromethane	5.379	113	49624	51.693	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.380%	
50) Toluene-d8	8.647	98	174614	52.113	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.220%	
62) 4-Bromofluorobenzene	11.079	95	64159	52.568	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.140%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	45908	88.875	ug/l	97
18) Methyl Acetate	2.703	43	2417	2.032	ug/l	92
20) Methylene Chloride	2.782	84	28428	29.397	ug/l	98
25) 2-Butanone	4.593	43	4497	5.948	ug/l	97
43) Isopropyl Acetate	6.349	43	7539	3.045	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045605.D
Acq On : 04 Apr 2025 17:29
Operator : JC/MD
Sample : PB167448ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167448ZHE#04

Quant Time: Apr 04 23:13:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

