

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
 Data File : VX045607.D
 Acq On : 04 Apr 2025 18:15
 Operator : JC/MD
 Sample : Q1694-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-18

Quant Time: Apr 04 23:14:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	63724	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	124983	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	117651	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	46504	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64231	55.118	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.240%
35) Dibromofluoromethane	5.379	113	45598	51.421	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.840%
50) Toluene-d8	8.647	98	162751	52.583	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.160%
62) 4-Bromofluorobenzene	11.079	95	56840	50.417	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.840%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	37636	78.650	ug/l	97
18) Methyl Acetate	2.709	43	2313	2.099	ug/l	92
20) Methylene Chloride	2.782	84	9017	10.065	ug/l	93
25) 2-Butanone	4.580	43	5931	8.467	ug/l	96
43) Isopropyl Acetate	6.348	43	7268	3.178	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045607.D
Acq On : 04 Apr 2025 18:15
Operator : JC/MD
Sample : Q1694-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-18

Quant Time: Apr 04 23:14:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

