

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
 Data File : VX045608.D
 Acq On : 04 Apr 2025 18:38
 Operator : JC/MD
 Sample : Q1695-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-2

Quant Time: Apr 04 23:14:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	67319	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	133833	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	126923	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53869	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67401	54.749	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.500%
35) Dibromofluoromethane	5.379	113	48597	51.179	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.360%
50) Toluene-d8	8.647	98	171101	51.625	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.260%
62) 4-Bromofluorobenzene	11.079	95	63473	52.577	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.160%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	37842	74.858	ug/l	97
18) Methyl Acetate	2.696	43	2171	1.865	ug/l	94
20) Methylene Chloride	2.782	84	8073	8.530	ug/l	95
25) 2-Butanone	4.580	43	5350	7.230	ug/l #	78
43) Isopropyl Acetate	6.348	43	6695	2.734	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045608.D
Acq On : 04 Apr 2025 18:38
Operator : JC/MD
Sample : Q1695-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-2

Quant Time: Apr 04 23:14:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

