

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
 Data File : VX045609.D
 Acq On : 04 Apr 2025 19:01
 Operator : JC/MD
 Sample : Q1695-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-3

Quant Time: Apr 04 23:14:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64273	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	127859	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	122107	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52275	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64751	55.089	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 110.180%		
35) Dibromofluoromethane	5.379	113	46929	51.732	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 103.460%		
50) Toluene-d8	8.647	98	162903	51.448	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 102.900%		
62) 4-Bromofluorobenzene	11.079	95	59053	51.202	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	= 102.400%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	38246	79.242	ug/l	99
18) Methyl Acetate	2.703	43	1942	1.748	ug/l #	84
20) Methylene Chloride	2.788	84	8607	9.526	ug/l	97
25) 2-Butanone	4.580	43	5116	7.241	ug/l	91
43) Isopropyl Acetate	6.348	43	7096	3.033	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045609.D
Acq On : 04 Apr 2025 19:01
Operator : JC/MD
Sample : Q1695-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-3

Quant Time: Apr 04 23:14:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

