

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
 Data File : VX045610.D
 Acq On : 04 Apr 2025 19:24
 Operator : JC/MD
 Sample : Q1712-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z-05A

Quant Time: Apr 04 23:14:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66374	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	131446	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	125072	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51775	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	66314	54.633	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.260%	
35) Dibromofluoromethane	5.385	113	47961	51.427	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.860%	
50) Toluene-d8	8.647	98	168938	51.898	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.800%	
62) 4-Bromofluorobenzene	11.079	95	61959	52.255	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.520%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	41903	84.071	ug/l	96
18) Methyl Acetate	2.703	43	1777	1.549	ug/l #	84
20) Methylene Chloride	2.788	84	9178	9.836	ug/l	93
25) 2-Butanone	4.586	43	5041	6.909	ug/l	97
43) Isopropyl Acetate	6.348	43	7465	3.104	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045610.D
Acq On : 04 Apr 2025 19:24
Operator : JC/MD
Sample : Q1712-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z-05A

Quant Time: Apr 04 23:14:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

