

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
 Data File : VX045611.D
 Acq On : 04 Apr 2025 19:48
 Operator : JC/MD
 Sample : Q1712-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-7

Quant Time: Apr 04 23:14:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	67805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	134871	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	130302	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53641	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69730	56.235	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.460%
35) Dibromofluoromethane	5.379	113	49438	51.664	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.320%
50) Toluene-d8	8.647	98	173797	52.035	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.080%
62) 4-Bromofluorobenzene	11.079	95	64846	53.301	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.600%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	40642	79.820	ug/l	96
18) Methyl Acetate	2.709	43	1812	1.546	ug/l	91
20) Methylene Chloride	2.788	84	8744	9.173	ug/l	94
25) 2-Butanone	4.580	43	5513	7.397	ug/l	96
43) Isopropyl Acetate	6.354	43	7228	2.929	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040425\
Data File : VX045611.D
Acq On : 04 Apr 2025 19:48
Operator : JC/MD
Sample : Q1712-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-7

Quant Time: Apr 04 23:14:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

