

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
 Data File : VX008674.D
 Acq On : 05 Apr 2019 10:51
 Operator : JC/SP
 Sample : VX0405WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0405WBL01

Quant Time: Apr 05 13:48:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\624X032719W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Mar 27 12:01:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	29960	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	177398	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	151379	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.06	65	75637	30.18	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.60%
60) 4-Bromofluorobenzene	11.13	95	72585	27.52	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	91.73%
63) Toluene-d8	8.71	98	217264	30.24	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
Data File : VX008674.D
Acq On : 05 Apr 2019 10:51
Operator : JC/SP
Sample : VX0405WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0405WBL01

Quant Time: Apr 05 13:48:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\624X032719W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Mar 27 12:01:15 2019
Response via : Initial Calibration

