

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
 Data File : VX008692.D
 Acq On : 05 Apr 2019 19:19
 Operator : JC/SP
 Sample : K2274-08DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST006RW149-190402DL

Quant Time: Apr 06 03:10:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	204681	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	339259	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	286488	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	118022	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	139713	46.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.60%	
35) Dibromofluoromethane	5.50	113	102924	47.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.90%	
50) Toluene-d8	8.71	98	423100	48.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.98%	
62) 4-Bromofluorobenzene	11.13	95	139699	45.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.22%	
Target Compounds						
4) Vinyl Chloride	1.40	62	10649	3.344	ug/l	96
27) cis-1,2-Dichloroethene	4.60	96	130138	44.223	ug/l	89
32) 1,1,1-Trichloroethane	5.50	97	1103	0.296	ug/l #	41
39) Methylcyclohexane	7.46	83	1559	0.340	ug/l #	89
44) Trichloroethene	7.21	130	33798	13.135	ug/l	99
64) Tetrachloroethene	9.34	164	38282	18.036	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
Data File : VX008692.D
Acq On : 05 Apr 2019 19:19
Operator : JC/SP
Sample : K2274-08DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST006RW149-190402DL

Quant Time: Apr 06 03:10:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

