

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
 Data File : VX008695.D
 Acq On : 05 Apr 2019 20:29
 Operator : JC/SP
 Sample : K2274-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST006MW148-190402

Quant Time: Apr 06 03:18:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	212092	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	351871	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	306843	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	134213	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	143643	46.43	ug/l	0.00
Spiked Amount			Recovery	=	92.86%	
35) Dibromofluoromethane	5.50	113	106626	47.40	ug/l	0.00
Spiked Amount			Recovery	=	94.80%	
50) Toluene-d8	8.71	98	445653	49.75	ug/l	0.00
Spiked Amount			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	11.13	95	152410	47.98	ug/l	0.00
Spiked Amount			Recovery	=	95.96%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	147285	44.628	ug/l	97
21) trans-1,2-Dichloroethene	3.17	96	1760	0.664	ug/l #	73
24) 1,1-Dichloroethane	3.70	63	3660	0.683	ug/l #	97
27) cis-1,2-Dichloroethene	4.60	96	141694	46.468	ug/l	90
31) Cyclohexane	5.58	56	18308	3.425	ug/l	97
32) 1,1,1-Trichloroethane	5.50	97	3148	0.814	ug/l #	52
39) Methylcyclohexane	7.46	83	42164	8.874	ug/l #	77
40) Benzene	6.14	78	10163	0.887	ug/l	99
67) Ethyl Benzene	10.25	91	6260	0.494	ug/l	94
68) m/p-Xylenes	10.36	106	1501	0.329	ug/l	93
73) Isopropylbenzene	11.02	105	4135	0.372	ug/l	96
83) tert-Butylbenzene	11.77	119	11670	1.299	ug/l	85
85) sec-Butylbenzene	11.94	105	22180	2.079	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
Data File : VX008695.D
Acq On : 05 Apr 2019 20:29
Operator : JC/SP
Sample : K2274-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST006MW148-190402

Quant Time: Apr 06 03:18:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

