

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
 Data File : VX008696.D
 Acq On : 05 Apr 2019 20:53
 Operator : JC/SP
 Sample : K2295-17DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW311-190404DL

Quant Time: Apr 06 03:22:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	211156	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	351985	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	311317	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	133757	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	144456	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
35) Dibromofluoromethane	5.50	113	106343	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
50) Toluene-d8	8.71	98	443048	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
62) 4-Bromofluorobenzene	11.13	95	154399	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
Target Compounds						
4) Vinyl Chloride	1.41	62	1838	0.559	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	75947	25.017	ug/l	88
44) Trichloroethene	7.21	130	15113	5.661	ug/l	97
64) Tetrachloroethene	9.34	164	21095	9.146	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
Data File : VX008696.D
Acq On : 05 Apr 2019 20:53
Operator : JC/SP
Sample : K2295-17DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW311-190404DL

Quant Time: Apr 06 03:22:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

