

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
 Data File : VX008697.D
 Acq On : 05 Apr 2019 21:16
 Operator : JC/SP
 Sample : K2295-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW025-190404

Quant Time: Apr 06 03:24:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	211779	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	352951	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	310957	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	131815	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	145676	47.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.32%	
35) Dibromofluoromethane	5.50	113	106927	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
50) Toluene-d8	8.71	98	444706	49.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.98%	
62) 4-Bromofluorobenzene	11.13	95	151421	47.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.04%	
Target Compounds						
4) Vinyl Chloride	1.41	62	14433	4.380	ug/l	100
27) cis-1,2-Dichloroethene	4.60	96	32641	10.720	ug/l	89
40) Benzene	6.15	78	4939	0.430	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
Data File : VX008697.D
Acq On : 05 Apr 2019 21:16
Operator : JC/SP
Sample : K2295-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW025-190404

Quant Time: Apr 06 03:24:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

