

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
 Data File : VX008702.D
 Acq On : 05 Apr 2019 23:12
 Operator : JC/SP
 Sample : K2296-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-TB-2019-2

Quant Time: Apr 06 03:37:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	207035	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	346579	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	299839	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	126813	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	144414	47.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.64%	
35) Dibromofluoromethane	5.50	113	104981	47.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.76%	
50) Toluene-d8	8.71	98	433349	49.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.24%	
62) 4-Bromofluorobenzene	11.13	95	146732	46.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.80%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
Data File : VX008702.D
Acq On : 05 Apr 2019 23:12
Operator : JC/SP
Sample : K2296-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-TB-2019-2

Quant Time: Apr 06 03:37:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

