

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
 Data File : VX008707.D
 Acq On : 06 Apr 2019 01:09
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 06 03:52:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.576	0.504	12.5	77	0.00
3 P	Chloromethane	0.795	0.710	10.7	86	0.00
4 C	Vinyl Chloride	0.778	0.703	9.6#	80	0.00
5 T	Bromomethane	0.415	0.316	23.9#	81	0.00
6 T	Chloroethane	0.516	0.422	18.2	81	0.00
7 T	Trichlorofluoromethane	0.860	0.791	8.0	81	0.00
8 T	Diethyl Ether	0.408	0.373	8.6	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.554	0.498	10.1	79	0.00
10 T	Methyl Iodide	0.456	0.450	1.3	76	0.00
11 T	Tert butyl alcohol	0.171	0.166	2.9	85	0.00
12 CM	1,1-Dichloroethene	0.578	0.521	9.9#	80	0.00
13 T	Acrolein	0.085	0.210	-147.1#	223#	0.00
14 T	Allyl chloride	1.323	1.213	8.3	80	0.00
15 T	Acrylonitrile	0.452	0.425	6.0	82	0.00
16 T	Acetone	0.420	0.364	13.3	72	0.00
17 T	Carbon Disulfide	1.857	1.500	19.2	76	0.00
18 T	Methyl Acetate	1.019	0.953	6.5	85	0.00
19 T	Methyl tert-butyl Ether	2.064	1.993	3.4	83	0.00
20 T	Methylene Chloride	0.703	0.633	10.0	83	0.00
21 T	trans-1,2-Dichloroethene	0.625	0.547	12.5	80	0.00
22 T	Diisopropyl ether	2.551	2.413	5.4	83	0.00
23 T	Vinyl Acetate	2.219	2.149	3.2	82	0.00
24 P	1,1-Dichloroethane	1.263	1.172	7.2	81	0.00
25 T	2-Butanone	0.664	0.632	4.8	80	0.00
26 T	2,2-Dichloropropane	0.889	0.715	19.6	69	0.00
27 T	cis-1,2-Dichloroethene	0.719	0.650	9.6	81	0.00
28 T	Bromochloromethane	0.599	0.551	8.0	83	0.00
29 T	Tetrahydrofuran	0.441	0.420	4.8	83	0.00
30 C	Chloroform	1.128	1.056	6.4#	82	0.00
31 T	Cyclohexane	1.260	1.129	10.4	81	0.00
32 T	1,1,1-Trichloroethane	0.911	0.864	5.2	81	0.00
33 S	1,2-Dichloroethane-d4	0.729	0.732	-0.4	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.320	0.321	-0.3	88	0.00
36 T	1,1-Dichloropropene	0.547	0.483	11.7	79	0.00
37 T	Ethyl Acetate	0.730	0.702	3.8	82	0.00
38 T	Carbon Tetrachloride	0.452	0.422	6.6	80	0.00
39 T	Methylcyclohexane	0.675	0.585	13.3	77	0.00
40 TM	Benzene	1.628	1.500	7.9	81	0.00
41 T	Methacrylonitrile	0.419	0.397	5.3	83	0.00
42 TM	1,2-Dichloroethane	0.584	0.534	8.6	82	0.00
43 T	Isopropyl Acetate	1.143	1.130	1.1	85	0.00
44 TM	Trichloroethene	0.379	0.344	9.2	80	0.00
45 C	1,2-Dichloropropane	0.465	0.428	8.0#	81	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
 Data File : VX008707.D
 Acq On : 06 Apr 2019 01:09
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 06 03:52:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.270	0.241	10.7	81	0.00
47 T	Bromodichloromethane	0.508	0.485	4.5	82	0.00
48 T	Methyl methacrylate	0.578	0.567	1.9	84	0.00
49 T	1,4-Dioxane	0.011	0.010	9.1	85	0.00
50 S	Toluene-d8	1.273	1.296	-1.8	88	0.00
51 T	4-Methyl-2-Pentanone	0.736	0.731	0.7	85	0.00
52 CM	Toluene	0.960	0.906	5.6#	82	0.00
53 T	t-1,3-Dichloropropene	0.569	0.547	3.9	79	0.00
54 T	cis-1,3-Dichloropropene	0.650	0.613	5.7	80	0.00
55 T	1,1,2-Trichloroethane	0.385	0.370	3.9	84	0.00
56 T	Ethyl methacrylate	0.697	0.703	-0.9	85	0.00
57 T	1,3-Dichloropropane	0.692	0.654	5.5	82	0.00
58 T	2-Chloroethyl Vinyl ether	0.356	0.346	2.8	83	0.00
59 T	2-Hexanone	0.572	0.549	4.0	83	0.00
60 T	Dibromochloromethane	0.363	0.361	0.6	83	0.00
61 T	1,2-Dibromoethane	0.401	0.380	5.2	82	0.00
62 S	4-Bromofluorobenzene	0.451	0.467	-3.5	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.370	0.334	9.7	82	0.00
65 PM	Chlorobenzene	1.108	1.020	7.9	84	0.00
66 T	1,1,1,2-Tetrachloroethane	0.389	0.366	5.9	83	0.00
67 C	Ethyl Benzene	2.067	1.903	7.9#	82	0.00
68 T	m/p-Xylenes	0.743	0.696	6.3	84	0.00
69 T	o-Xylene	0.720	0.676	6.1	84	0.00
70 T	Styrene	1.241	1.179	5.0	84	0.00
71 P	Bromoform	0.297	0.293	1.3	84	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	4.144	3.931	5.1	83	0.00
74 T	N-amyl acetate	2.438	2.335	4.2	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.576	1.445	8.3	83	0.00
76 T	1,2,3-Trichloropropane	1.374	1.424	-3.6	95	0.00
77 T	Bromobenzene	0.996	0.940	5.6	84	0.00
78 T	n-propylbenzene	4.825	4.555	5.6	82	0.00
79 T	2-Chlorotoluene	2.920	2.713	7.1	84	0.00
80 T	1,3,5-Trimethylbenzene	3.478	3.324	4.4	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.498	0.467	6.2	77	0.00
82 T	4-Chlorotoluene	3.380	3.142	7.0	83	0.00
83 T	tert-Butylbenzene	3.346	3.243	3.1	86	0.00
84 T	1,2,4-Trimethylbenzene	3.522	3.334	5.3	83	0.00
85 T	sec-Butylbenzene	3.975	3.771	5.1	82	0.00
86 T	p-Isopropyltoluene	3.568	3.377	5.4	82	0.00
87 T	1,3-Dichlorobenzene	1.783	1.630	8.6	83	0.00
88 T	1,4-Dichlorobenzene	1.806	1.645	8.9	84	0.00
89 T	n-Butylbenzene	3.346	3.120	6.8	80	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX040519\
Data File : VX008707.D
Acq On : 06 Apr 2019 01:09
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 06 03:52:44 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.592	0.574	3.0	83	0.00
91 T	1,2-Dichlorobenzene	1.771	1.638	7.5	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.347	0.326	6.1	85	0.00
93 T	1,2,4-Trichlorobenzene	1.231	1.137	7.6	84	0.00
94 T	Hexachlorobutadiene	0.590	0.552	6.4	82	0.00
95 T	Naphthalene	4.115	3.940	4.3	86	0.00
96 T	1,2,3-Trichlorobenzene	1.230	1.139	7.4	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6