

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040523\
 Data File : VX034941.D
 Acq On : 05 Apr 2023 10:10
 Operator : JC/MD
 Sample : VX0405MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0405MBL01

Quant Time: Apr 06 02:09:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	260941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	442619	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	397423	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	174824	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	197835	49.574	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.140%
35) Dibromofluoromethane	5.385	113	144600	49.392	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.780%
50) Toluene-d8	8.646	98	531061	49.146	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.300%
62) 4-Bromofluorobenzene	11.079	95	203161	48.585	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040523\
Data File : VX034941.D
Acq On : 05 Apr 2023 10:10
Operator : JC/MD
Sample : VX0405MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0405MBL01

Quant Time: Apr 06 02:09:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

