

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040523\
 Data File : VX034945.D
 Acq On : 05 Apr 2023 11:50
 Operator : JC/MD
 Sample : 02190-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SY-3

Quant Time: Apr 06 02:12:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	258296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	433266	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	377790	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	160113	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	189130	47.878	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.760%
35) Dibromofluoromethane	5.385	113	139221	48.582	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.160%
50) Toluene-d8	8.647	98	517637	48.938	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.880%
62) 4-Bromofluorobenzene	11.079	95	188897	46.149	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.300%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	4580	2.854	ug/l	92
44) Trichloroethene	7.141	130	1170	0.376	ug/l	71
89) n-Butylbenzene	12.335	91	2459	0.301	ug/l	87
94) Hexachlorobutadiene	13.719	225	1271	0.964	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040523\
Data File : VX034945.D
Acq On : 05 Apr 2023 11:50
Operator : JC/MD
Sample : 02190-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SY-3

Quant Time: Apr 06 02:12:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

