

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040524\
 Data File : VX040892.D
 Acq On : 05 Apr 2024 10:50
 Operator : JC/MD
 Sample : PB160019TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB160019TB

Quant Time: Apr 05 23:17:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	44484	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	81636	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	83422	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	43708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	39310	49.250	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.500%
35) Dibromofluoromethane	5.385	113	28551	49.377	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.760%
50) Toluene-d8	8.647	98	105302	50.398	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.800%
62) 4-Bromofluorobenzene	11.079	95	43349	51.952	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.900%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	15760	50.852	ug/l	98
20) Methylene Chloride	2.794	84	5490	9.559	ug/l	91
43) Isopropyl Acetate	6.336	43	39677	25.466	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040524\
Data File : VX040892.D
Acq On : 05 Apr 2024 10:50
Operator : JC/MD
Sample : PB160019TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB160019TB

Quant Time: Apr 05 23:17:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

