

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040524\
 Data File : VX040893.D
 Acq On : 05 Apr 2024 11:13
 Operator : JC/MD
 Sample : PB160019ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB160019ZHE#01

Quant Time: Apr 05 23:17:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	44697	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	80689	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	84173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	43818	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	39205	48.885	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.760%
35) Dibromofluoromethane	5.385	113	27864	48.754	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.500%
50) Toluene-d8	8.647	98	105290	50.984	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.960%
62) 4-Bromofluorobenzene	11.079	95	43680	52.964	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.920%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	11197	35.956	ug/l	98
20) Methylene Chloride	2.794	84	10664	18.479	ug/l	93
43) Isopropyl Acetate	6.336	43	37899	24.610	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040524\
Data File : VX040893.D
Acq On : 05 Apr 2024 11:13
Operator : JC/MD
Sample : PB160019ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB160019ZHE#01

Quant Time: Apr 05 23:17:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

