

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040524\
 Data File : VX040894.D
 Acq On : 05 Apr 2024 11:36
 Operator : JC/MD
 Sample : PB160019ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB160019ZHE#02

Quant Time: Apr 05 23:18:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	44385	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	81161	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	83964	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	44075	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	39174	49.189	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.380%
35) Dibromofluoromethane	5.391	113	28170	49.003	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.000%
50) Toluene-d8	8.647	98	104023	50.077	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.160%
62) 4-Bromofluorobenzene	11.079	95	44210	53.294	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.580%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	10404	33.645	ug/l	93
20) Methylene Chloride	2.788	84	10331	18.028	ug/l	99
43) Isopropyl Acetate	6.336	43	37058	23.924	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040524\
Data File : VX040894.D
Acq On : 05 Apr 2024 11:36
Operator : JC/MD
Sample : PB160019ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB160019ZHE#02

Quant Time: Apr 05 23:18:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

