

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040524\
 Data File : VX040901.D
 Acq On : 05 Apr 2024 14:16
 Operator : JC/MD
 Sample : P2002-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-040424

Quant Time: Apr 05 23:21:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	41401	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	77131	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	76001	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	39150	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	38144	51.348	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.700%
35) Dibromofluoromethane	5.385	113	27135	49.669	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.340%
50) Toluene-d8	8.647	98	96515	48.890	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.780%
62) 4-Bromofluorobenzene	11.079	95	39993	50.730	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.460%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	1289	2.610	ug/l	99
16) Acetone	2.379	43	2510	8.702	ug/l	94
20) Methylene Chloride	2.794	84	1448	2.709	ug/l #	93
44) Trichloroethene	7.123	130	31149	57.887	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040524\
Data File : VX040901.D
Acq On : 05 Apr 2024 14:16
Operator : JC/MD
Sample : P2002-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-040424

Quant Time: Apr 05 23:21:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

