

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040524\
 Data File : VX040902.D
 Acq On : 05 Apr 2024 14:39
 Operator : JC/MD
 Sample : P2002-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RMW-06B-44-040424

Quant Time: Apr 05 23:21:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	38635	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	70090	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	71000	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38034	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	34972	50.449	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.900%	
35) Dibromofluoromethane	5.379	113	24511	49.373	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.740%	
50) Toluene-d8	8.647	98	89030	49.629	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.260%	
62) 4-Bromofluorobenzene	11.079	95	37731	52.669	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.340%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	739	1.604	ug/l	94
16) Acetone	2.380	43	995	3.697	ug/l #	77
27) cis-1,2-Dichloroethene	4.489	96	1183	2.244	ug/l	83
30) Chloroform	5.093	83	3104	3.313	ug/l	91
44) Trichloroethene	7.123	130	20786	42.509	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040524\
Data File : VX040902.D
Acq On : 05 Apr 2024 14:39
Operator : JC/MD
Sample : P2002-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RMW-06B-44-040424

Quant Time: Apr 05 23:21:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

