

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040524\
 Data File : VX040909.D
 Acq On : 05 Apr 2024 17:17
 Operator : JC/MD
 Sample : P1996-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP1

Quant Time: Apr 05 23:24:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 02 05:40:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	41288	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	78063	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	77684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	40106	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	37839	51.077	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.160%	
35) Dibromofluoromethane	5.385	113	26324	47.609	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.220%	
50) Toluene-d8	8.647	98	97977	49.038	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.080%	
62) 4-Bromofluorobenzene	11.079	95	41072	51.477	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.960%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	24695	85.850	ug/l	99
20) Methylene Chloride	2.794	84	2498	4.686	ug/l	97
25) 2-Butanone	4.568	43	2424	5.673	ug/l	100
43) Isopropyl Acetate	6.336	43	44644	29.965	ug/l	97
44) Trichloroethene	7.129	130	1742	3.199	ug/l	96
52) Toluene	8.720	92	2003	1.456	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040524\
Data File : VX040909.D
Acq On : 05 Apr 2024 17:17
Operator : JC/MD
Sample : P1996-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP1

Quant Time: Apr 05 23:24:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 02 05:40:55 2024
Response via : Initial Calibration

