



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040524\
Data File : VX040886.D
Acq On : 05 Apr 2024 08:21
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040124W.M
Title : SW846 8260
Last Update : Tue Apr 02 05:40:55 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	3567	PASS
75	95	30	60	58.2	9200	PASS
95	95	100	100	100.0	15819	PASS
96	95	5	9	6.5	1025	PASS
173	174	0.00	2	1.2	133	PASS
174	95	50	100	71.7	11346	PASS
175	174	5	9	7.7	870	PASS
176	174	95	101	98.4	11159	PASS
177	176	5	9	6.4	715	PASS