

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040619\
 Data File : VX008714.D
 Acq On : 06 Apr 2019 12:45
 Operator : JC/SP
 Sample : K2292-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Apr 15 02:15:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	209123	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	345528	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	302939	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	129468	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	141925	46.53	ug/l	0.00
Spiked Amount						
			Recovery	=	93.06%	
35) Dibromofluoromethane	5.50	113	104758	47.42	ug/l	0.00
Spiked Amount						
			Recovery	=	94.84%	
50) Toluene-d8	8.71	98	437241	49.71	ug/l	0.00
Spiked Amount						
			Recovery	=	99.42%	
62) 4-Bromofluorobenzene	11.13	95	150018	48.09	ug/l	0.00
Spiked Amount						
			Recovery	=	96.18%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040619\
Data File : VX008714.D
Acq On : 06 Apr 2019 12:45
Operator : JC/SP
Sample : K2292-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Apr 15 02:15:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

