

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040619\
 Data File : VX008723.D
 Acq On : 06 Apr 2019 16:15
 Operator : JC/SP
 Sample : K2295-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW026-190404

Quant Time: Apr 08 03:59:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	209505	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	348630	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	303116	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	126605	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	143436	46.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.88%	
35) Dibromofluoromethane	5.50	113	105165	47.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.36%	
50) Toluene-d8	8.71	98	436654	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
62) 4-Bromofluorobenzene	11.13	95	147192	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	69684	23.135	ug/l	89
44) Trichloroethene	7.21	130	15375	5.815	ug/l	92
64) Tetrachloroethene	9.34	164	151066	67.267	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040619\
Data File : VX008723.D
Acq On : 06 Apr 2019 16:15
Operator : JC/SP
Sample : K2295-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW026-190404

Quant Time: Apr 08 03:59:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

