

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040619\
 Data File : VX008724.D
 Acq On : 06 Apr 2019 16:38
 Operator : JC/SP
 Sample : K2295-03 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW026-190404

Quant Time: Apr 08 04:10:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	202674	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	333431	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	288381	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	120171	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	137627	46.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.12%	
35) Dibromofluoromethane	5.50	113	100199	47.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.00%	
50) Toluene-d8	8.71	98	416655	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
62) 4-Bromofluorobenzene	11.13	95	139260	46.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.52%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	7725	2.651	ug/l	89
44) Trichloroethene	7.21	130	1984	0.785	ug/l	88
64) Tetrachloroethene	9.34	164	19305	9.035	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040619\
Data File : VX008724.D
Acq On : 06 Apr 2019 16:38
Operator : JC/SP
Sample : K2295-03 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW026-190404

Quant Time: Apr 08 04:10:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

