

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040619\
 Data File : VX008728.D
 Acq On : 06 Apr 2019 18:11
 Operator : JC/SP
 Sample : K2295-06 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008RW256-190404

Quant Time: Apr 08 04:21:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	205191	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	338697	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	288859	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	117063	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	139965	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
35) Dibromofluoromethane	5.50	113	102242	47.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.42%	
50) Toluene-d8	8.71	98	425035	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
62) 4-Bromofluorobenzene	11.13	95	137024	44.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.62%	
Target Compounds						
4) Vinyl Chloride	1.40	62	1491	0.467	ug/l	93
27) cis-1,2-Dichloroethene	4.60	96	8595	2.913	ug/l	92
44) Trichloroethene	7.21	130	2669	1.039	ug/l	99
64) Tetrachloroethene	9.34	164	5353	2.501	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040619\
Data File : VX008728.D
Acq On : 06 Apr 2019 18:11
Operator : JC/SP
Sample : K2295-06 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008RW256-190404

Quant Time: Apr 08 04:21:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

