

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040619\
 Data File : VX008729.D
 Acq On : 06 Apr 2019 18:35
 Operator : JC/SP
 Sample : K2295-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW107-190403

Quant Time: Apr 08 04:24:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	202050	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	332736	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	292930	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	124535	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	138074	46.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.70%	
35) Dibromofluoromethane	5.50	113	100454	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
50) Toluene-d8	8.71	98	419094	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
62) 4-Bromofluorobenzene	11.13	95	143493	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
Target Compounds						
4) Vinyl Chloride	1.40	62	5990	1.905	ug/l	92
27) cis-1,2-Dichloroethene	4.60	96	14793	5.092	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040619\
Data File : VX008729.D
Acq On : 06 Apr 2019 18:35
Operator : JC/SP
Sample : K2295-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW107-190403

Quant Time: Apr 08 04:24:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

