

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040619\
 Data File : VX008733.D
 Acq On : 06 Apr 2019 20:08
 Operator : JC/SP
 Sample : K2295-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-TB04-190403

Quant Time: Apr 08 04:40:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 04 04:47:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	211425	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	350729	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	302926	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	125398	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	143157	46.42	ug/l	0.00
Spiked Amount			Recovery	=	92.84%	
35) Dibromofluoromethane	5.50	113	105564	47.08	ug/l	0.00
Spiked Amount			Recovery	=	94.16%	
50) Toluene-d8	8.71	98	439866	49.26	ug/l	0.00
Spiked Amount			Recovery	=	98.52%	
62) 4-Bromofluorobenzene	11.13	95	144896	45.76	ug/l	0.00
Spiked Amount			Recovery	=	91.52%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040619\
Data File : VX008733.D
Acq On : 06 Apr 2019 20:08
Operator : JC/SP
Sample : K2295-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-TB04-190403

Quant Time: Apr 08 04:40:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X040319W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 04 04:47:41 2019
Response via : Initial Calibration

