

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040620\
 Data File : VX015508.D
 Acq On : 06 Apr 2020 13:33
 Operator : JC/SP
 Sample : L2097-15 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40788

Quant Time: Apr 07 06:10:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1153249	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1857250	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1823400	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	987935	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	777418	47.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.60%	
35) Dibromofluoromethane	5.47	113	584947	49.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.54%	
50) Toluene-d8	8.70	98	2265639	51.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.66%	
62) 4-Bromofluorobenzene	11.12	95	988838	55.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.34%	
Target Compounds						
16) Acetone	2.42	43	178594	25.557	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040620\
Data File : VX015508.D
Acq On : 06 Apr 2020 13:33
Operator : JC/SP
Sample : L2097-15 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40788

Quant Time: Apr 07 06:10:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

