

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040620\
 Data File : VX015520.D
 Acq On : 06 Apr 2020 18:04
 Operator : JC/SP
 Sample : L2112-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HR-06-040220

Quant Time: Apr 07 07:21:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1139865	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1876524	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1792697	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	1018838	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	780135	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
35) Dibromofluoromethane	5.47	113	598037	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
50) Toluene-d8	8.70	98	2271569	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
62) 4-Bromofluorobenzene	11.12	95	996639	55.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.06%	
Target Compounds						
16) Acetone	2.42	43	63124	9.139	ug/l	97
20) Methylene Chloride	2.84	84	37175	2.882	ug/l	99
43) Isopropyl Acetate	6.42	43	387403	10.194	ug/l	99
84) 1,2,4-Trimethylbenzene	11.79	105	107703	1.854	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040620\
Data File : VX015520.D
Acq On : 06 Apr 2020 18:04
Operator : JC/SP
Sample : L2112-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HR-06-040220

Quant Time: Apr 07 07:21:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

