

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040620\
 Data File : VX015524.D
 Acq On : 06 Apr 2020 19:35
 Operator : JC/SP
 Sample : L2164-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ROLL-OFF-COMP

Quant Time: Apr 07 07:36:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1161854	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1869156	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1773990	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	996818	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	768814	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
35) Dibromofluoromethane	5.47	113	586779	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
50) Toluene-d8	8.70	98	2235055	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
62) 4-Bromofluorobenzene	11.12	95	987678	55.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.50%	
Target Compounds						
16) Acetone	2.42	43	64394	9.147	ug/l	97
20) Methylene Chloride	2.84	84	21741	1.653	ug/l	94
43) Isopropyl Acetate	6.42	43	398255	10.521	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040620\
Data File : VX015524.D
Acq On : 06 Apr 2020 19:35
Operator : JC/SP
Sample : L2164-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ROLL-OFF-COMP

Quant Time: Apr 07 07:36:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

