

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040620\
 Data File : VX015525.D
 Acq On : 06 Apr 2020 19:57
 Operator : JC/SP
 Sample : L2097-15 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40788

Quant Time: Apr 07 07:40:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1152589	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1849726	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1786450	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	1031072	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	758762	46.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.36%	
35) Dibromofluoromethane	5.47	113	581580	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
50) Toluene-d8	8.70	98	2235448	51.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.70%	
62) 4-Bromofluorobenzene	11.12	95	999294	56.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.96%	
Target Compounds						
11) Tert butyl alcohol	3.00	59	58028	14.914	ug/l	97
16) Acetone	2.42	43	1007560	144.268	ug/l	99
25) 2-Butanone	4.64	43	101492	9.036	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040620\
Data File : VX015525.D
Acq On : 06 Apr 2020 19:57
Operator : JC/SP
Sample : L2097-15 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40788

Quant Time: Apr 07 07:40:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

