

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX040620\
 Data File : VX015526.D
 Acq On : 06 Apr 2020 20:20
 Operator : JC/SP
 Sample : L2097-16DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40851DL

Quant Time: Apr 07 07:49:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1148647	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1829467	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1810890	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	993880	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	766131	47.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.58%	
35) Dibromofluoromethane	5.47	113	586370	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
50) Toluene-d8	8.70	98	2245993	52.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.32%	
62) 4-Bromofluorobenzene	11.12	95	989888	56.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.14%	
Target Compounds						
16) Acetone	2.42	43	31837	4.574	ug/l	94
30) Chloroform	5.18	83	1148254	50.485	ug/l	97
47) Bromodichloromethane	7.88	83	135495	7.176	ug/l	99
60) Dibromochloromethane	9.57	129	56519	3.771	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX040620\
Data File : VX015526.D
Acq On : 06 Apr 2020 20:20
Operator : JC/SP
Sample : L2097-16DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40851DL

Quant Time: Apr 07 07:49:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

