

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX040620\
 Data File : VX015501.D
 Acq On : 06 Apr 2020 10:24
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 07 05:45:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	131	0.00
2 T	Dichlorodifluoromethane	0.441	0.447	-1.4	164#	0.00
3 P	Chloromethane	0.561	0.507	9.6	110	0.00
4 C	Vinyl Chloride	0.515	0.496	3.7#	133	0.00
5 T	Bromomethane	0.259	0.224	13.5	110	0.00
6 T	Chloroethane	0.319	0.305	4.4	125	0.00
7 T	Trichlorofluoromethane	0.691	0.709	-2.6	158#	0.00
8 T	Diethyl Ether	0.291	0.271	6.9	111	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.450	0.465	-3.3	168#	0.00
10 T	Methyl Iodide	0.607	0.537	11.5	97	0.00
11 T	Tert butyl alcohol	0.169	0.128	24.3#	89	0.00
12 CM	1,1-Dichloroethene	0.472	0.450	4.7#	130	0.00
13 T	Acrolein	0.096	0.088	8.3	109	0.00
14 T	Allyl chloride	1.103	1.026	7.0	112	0.00
15 T	Acrylonitrile	0.322	0.275	14.6	95	0.00
16 T	Acetone	0.303	0.274	9.6	103	0.00
17 T	Carbon Disulfide	1.440	1.330	7.6	125	0.00
18 T	Methyl Acetate	0.773	0.664	14.1	96	0.00
19 T	Methyl tert-butyl Ether	1.831	1.696	7.4	105	0.00
20 T	Methylene Chloride	0.566	0.510	9.9	108	0.00
21 T	trans-1,2-Dichloroethene	0.524	0.491	6.3	118	0.00
22 T	Diisopropyl ether	1.954	1.856	5.0	109	0.00
23 T	Vinyl Acetate	1.770	1.625	8.2	102	0.00
24 P	1,1-Dichloroethane	1.029	0.955	7.2	114	0.00
25 T	2-Butanone	0.487	0.412	15.4	95	0.00
26 T	2,2-Dichloropropane	0.845	0.898	-6.3	139	0.00
27 T	cis-1,2-Dichloroethene	0.606	0.567	6.4	111	0.00
28 T	Bromochloromethane	0.491	0.444	9.6	103	0.00
29 T	Tetrahydrofuran	0.313	0.257	17.9	92	0.00
30 C	Chloroform	0.990	0.931	6.0#	112	-0.01
31 T	Cyclohexane	0.893	0.896	-0.3	159#	0.00
32 T	1,1,1-Trichloroethane	0.896	0.869	3.0	127	0.00
33 S	1,2-Dichloroethane-d4	0.705	0.626	11.2	106	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	110	0.00
35 S	Dibromofluoromethane	0.320	0.311	2.8	108	-0.01
36 T	1,1-Dichloropropene	0.461	0.485	-5.2	132	0.00
37 T	Ethyl Acetate	0.600	0.548	8.7	94	-0.01
38 T	Carbon Tetrachloride	0.499	0.526	-5.4	135	0.00
39 T	Methylcyclohexane	0.531	0.598	-12.6	174#	0.00
40 TM	Benzene	1.350	1.371	-1.6	112	0.00
41 T	Methacrylonitrile	0.333	0.317	4.8	96	0.00
42 TM	1,2-Dichloroethane	0.542	0.549	-1.3	106	0.00
43 T	Isopropyl Acetate	1.013	0.966	4.6	98	0.00
44 TM	Trichloroethene	0.368	0.389	-5.7	124	0.00
45 C	1,2-Dichloropropane	0.369	0.371	-0.5#	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX040620\
 Data File : VX015501.D
 Acq On : 06 Apr 2020 10:24
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 07 05:45:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.242	0.242	0.0	105	0.00
47 T	Bromodichloromethane	0.516	0.531	-2.9	108	0.00
48 T	Methyl methacrylate	0.506	0.481	4.9	98	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	96	0.00
50 S	Toluene-d8	1.177	1.156	1.8	113	0.00
51 T	4-Methyl-2-Pentanone	0.574	0.537	6.4	95	0.00
52 CM	Toluene	0.832	0.878	-5.5#	116	0.00
53 T	t-1,3-Dichloropropene	0.605	0.617	-2.0	108	0.00
54 T	cis-1,3-Dichloropropene	0.616	0.641	-4.1	110	0.00
55 T	1,1,2-Trichloroethane	0.341	0.345	-1.2	106	0.00
56 T	Ethyl methacrylate	0.601	0.594	1.2	102	0.00
57 T	1,3-Dichloropropane	0.614	0.603	1.8	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.292	0.291	0.3	102	0.00
59 T	2-Hexanone	0.433	0.427	1.4	99	0.00
60 T	Dibromochloromethane	0.410	0.416	-1.5	104	0.00
61 T	1,2-Dibromoethane	0.371	0.372	-0.3	104	0.00
62 S	4-Bromofluorobenzene	0.478	0.478	0.0	114	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.377	0.415	-10.1	135	0.00
65 PM	Chlorobenzene	0.997	1.047	-5.0	115	0.00
66 T	1,1,1,2-Tetrachloroethane	0.398	0.408	-2.5	109	0.00
67 C	Ethyl Benzene	1.769	1.879	-6.2#	120	0.00
68 T	m/p-Xylenes	0.659	0.700	-6.2	120	0.00
69 T	o-Xylene	0.651	0.678	-4.1	114	0.00
70 T	Styrene	1.145	1.189	-3.8	113	0.00
71 P	Bromoform	0.363	0.364	-0.3	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
73 T	Isopropylbenzene	3.276	3.395	-3.6	125	0.00
74 T	N-amyl acetate	1.822	1.739	4.6	102	0.00
75 P	1,1,2,2-Tetrachloroethane	1.124	1.053	6.3	102	0.00
76 T	1,2,3-Trichloropropane	1.028	1.020	0.8	104	0.00
77 T	Bromobenzene	0.874	0.920	-5.3	116	0.00
78 T	n-propylbenzene	3.746	4.031	-7.6	129	0.00
79 T	2-Chlorotoluene	2.296	2.383	-3.8	118	0.00
80 T	1,3,5-Trimethylbenzene	2.798	2.988	-6.8	125	0.00
81 T	trans-1,4-Dichloro-2-butene	0.429	0.419	2.3	102	0.00
82 T	4-Chlorotoluene	2.692	2.843	-5.6	120	0.00
83 T	tert-Butylbenzene	2.733	2.930	-7.2	129	0.00
84 T	1,2,4-Trimethylbenzene	2.851	2.989	-4.8	120	0.00
85 T	sec-Butylbenzene	3.161	3.465	-9.6	139	0.00
86 T	p-Isopropyltoluene	2.934	3.269	-11.4	136	0.00
87 T	1,3-Dichlorobenzene	1.564	1.631	-4.3	118	0.00
88 T	1,4-Dichlorobenzene	1.593	1.652	-3.7	117	0.00
89 T	n-Butylbenzene	2.652	2.991	-12.8	141	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX040620\
Data File : VX015501.D
Acq On : 06 Apr 2020 10:24
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 07 05:45:07 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.571	0.596	-4.4	125	0.00
91 T	1,2-Dichlorobenzene	1.531	1.607	-5.0	116	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.317	0.276	12.9	99	0.00
93 T	1,2,4-Trichlorobenzene	1.211	1.310	-8.2	124	0.00
94 T	Hexachlorobutadiene	0.577	0.645	-11.8	154#	0.00
95 T	Naphthalene	3.694	3.710	-0.4	107	0.00
96 T	1,2,3-Trichlorobenzene	1.166	1.295	-11.1	121	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6